



ANDREA GALLO DI LUIGI S.r.l.u.

Azienda fondata nel 1892

www.andreagallo.it

Certificate of Analysis

M1QC007Rev.04

WATER bi-DISTILLED

COD. 04287 C

	BATCH Nr Typical	GRADE PG	RE-TEST DATE After 18 months
TEST	RESULTS	SPECS	PACKING
APPEARANCE of SOLUTION		Clear and Colourless	
CONDUCTIVITY (at 25 ° C)		Liquid 5.0	µS/cm max
ELECTRICAL RESISTIVITY (at 25 ° C)		0.2	MΩ cm min
pH (at 25 ° C)		5.0 - 8.0	
SODIUM (Na)		50	µg/L max
CHLORIDES (Cl)		50	µg/L max
			MANUFACTURING DATE

REMARKS :

- PG

- This document has been produced electronically and is valid without a signature
- Certified true and correct
- Issued by QC
- according to ASTM D1193-06 Type IV
- according to ASTM D1125